

Directions for Use

Primer for Resin Modified Glass Ionomer Cement

Oxford GI Resin PRIME VLC is a simple to use primer for dentine and enamel to enhance the bond between light cure resin modified glassionomer cements (e.g. Oxford GI Resin FILL UF) and tooth structure.

Contraindications

Do not use Oxford GI Resin PRIME VLC if the recommended working technique is not possible or if the patient is known to be allergic to any of the ingredients.

Side effects

In singular cases, Oxford GI Resin PRIME VLC may cause a sensitizing reaction in patients with a hypersensitivity to any of the ingredients. In these cases, the material should not be used.

Irritations resulting from direct contact with the pulp cannot be ruled out. Therefore for pulp protection the cavity floor in cases of deep excavations should be covered with a thin layer of calcium hydroxide material.

Incompatibility with Other Materials

Do not use in combination with substances containing eugenol because eugenol inhibits the polymerization of Oxford GI Resin PRIME VLC. Neither store the material in proximity of eugenol containing products, nor let the material allow coming into contact with materials containing eugenol.

Application

1. Isolation

Rubber dam is the recommended method of isolation.

2. Cavity Preparation

Clean the tooth with flour of pumice and water prior to preparation. Prepare the cavity with minimal tooth reduction. Margins should have a slight (0.5 - 1.0 mm) bevel placed in the enamel to increase the surface area for greater bond strength.

3. Pulp Protection

Cavity floor of deep excavations should be covered with a thin layer of calcium hydroxide material.

4. Application of Oxford GI Resin PRIME VLC

Apply Oxford GI Resin PRIME VLC generously with a brush onto the moist enamel and dentine surfaces for 30 seconds with agitation.

The material should build a homogeneous layer. Air thin for 10 seconds to remove the volatile components and to disperse the adhesive.

Then light cure for 20 sec with a dental halogen light unit or an LED (wavelength 400–500 nm, light intensity min. 1000 mW/cm²) and place the light cure resin reinforced glassionomer cement.

5. Restorative Placement

Refer to manufacturers instructions for placement, curing and finishing of light cure resin reinforced glassionomer cements.

Warnings

- Unpolymerized material may have an irritating effect and may lead to a sensitizing reaction against methacrylates
- Avoid contact with skin, mucous membrane and eyes
- If the material comes into contact with skin, immediately wash with water and soap. If the material comes into contact with eyes, immediately rinse with copious amounts of water and seek medical advice if required.
- Commercial medical gloves do not protect against the sensitizing effect of methacrylates.

Storage

Store between 4 and 25 °C (39 - 77 °F)!

Avoid storage in direct sunlight.

Do not use after expiration date (see expiration date on label/packaging)

Warranty

First Scientific Dental Materials GmbH warrants this product will be free from defects in material and manufacture. First Scientific Dental Materials makes no other warranties including any implied warranty of merchantability or fitness for a particular purpose. User is responsible for determining the suitability of the product for user's application. If this product is defective within the warranty period, your exclusively remedy and First Scientific Dental Materials' sole obligation shall be repair or replacement of the First Scientific Dental Materials product.

Limitation of Liability

Except where prohibited by law, First Scientific Dental Materials GmbH will not be liable for any loss or damage arising from this product, whether direct, indirect, special, incidental or consequential, regardless of the theory asserted, including warranty, contract, negligence or strict liability.

Keep away from children!

For dental use only!

Caution:

Federal law restricts the sale of this device to or by the order of a dentist.

